

PURPLE NITRILE-XTRA*

Exam Gloves

Product Specifications

Gauge Thickness Measurements	MM	MIL
Middle Finger:	.15	5.9
Palm:	.12	4.7
Cuff:	.11	4.3
Average Length	310mm	(12")

Physical Properties

	After Aging
Tensile Strength:	21 MPa
Ultimate Elongation:	550%

Quality Standards

Meets or exceeds ASTM D6319 - Standard Specification for Nitrile Examination Glove for Medical Applications.

Biocompatible per ISO 10993 studies for systemic toxicity, irritation and sensitization endpoints.

Passes Viral Penetration Testing - ASTM F1671 - 07 Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System.

Design Enhancements



Not formulated with Natural Rubber Latex – primary material is acrylonitrile butadiene, reducing the potential for glove-associated Type I allergic reactions to natural rubber latex proteins.



Powder-free, minimizing the potential for powder-related complications, such as irritant contact dermatitis.

All physical properties are target measurements; target MIL measurements may differ due to rounding.

HALYARD* PURPLE NITRILE-XTRA* Exam Gloves

PURPLE Nitrile-XTRA* Exam Gloves have an AQL of 1.0 for critical defects. This means that PURPLE Nitrile-XTRA* Exam Gloves are significantly less likely to have critical defects than exam gloves that just meet the ASTM 2.5 AQL requirement.

Sterile

Item	Description	Size	Packaging
14259	PURPLE NITRILE-XTRA* Sterile Pairs Exam Gloves	X-Small	50 pairs/box 4 boxes/case
14260	PURPLE NITRILE-XTRA* Sterile Pairs Exam Gloves	Small	50 pairs/box 4 boxes/case
14261	PURPLE NITRILE-XTRA* Sterile Pairs Exam Gloves	Medium	50 pairs/box 4 boxes/case
14262	PURPLE NITRILE-XTRA* Sterile Pairs Exam Gloves	Large	50 pairs/box 4 boxes/case
14263	PURPLE NITRILE-XTRA* Sterile Pairs Exam Gloves	X-Large	50 pairs/box 4 boxes/case



Distributed by Head to Heels Safety Supplies
9021 SW 94th Street, Unit 510, Miami FL 33176.
Ph: 305-712-6653
Email: cservice@headtoheels.net